

**Acceptance protocol**

SN:   
Contrast: 20:1  
Date:   
Operator:

**Optical density (OD) values gray patches\***

**Patch      Reference OD    Measured OD    Deviation**

1 (white)	0,10		
2	0,18		
3	0,26		
4	0,35		
5	0,44		
6	0,54		
7	0,65		
8	0,77		
9	0,90		
10	1,05		
11	1,21		
12 (black)	1,40		
BG	0,54		

\*Measured with X-Rite eXact

[https://www.image-engineering.de/content/products/charts/IE\\_reference\\_data\\_accuracy.pdf](https://www.image-engineering.de/content/products/charts/IE_reference_data_accuracy.pdf)

**Acceptance protocol**

SN:   
Contrast: 80:1  
Date:   
Operator:

**Optical density (OD) values gray patches\***

**Patch      Reference OD    Measured OD    Deviation**

Patch	Reference OD	Measured OD	Deviation
1 (white)	0,10		
2	0,19		
3	0,30		
4	0,41		
5	0,53		
6	0,66		
7	0,81		
8	0,97		
9	1,16		
10	1,39		
11	1,66		
12 (black)	2,00		
BG	1,11		

\*Measured with X-Rite eXact

[https://www.image-engineering.de/content/products/charts/IE\\_reference\\_data\\_accuracy.pdf](https://www.image-engineering.de/content/products/charts/IE_reference_data_accuracy.pdf)

**Acceptance protocol**

SN:   
Contrast: 160:1  
Date:   
Operator:

**Optical density (OD) values gray patches\***

**Patch      Reference OD    Measured OD    Deviation**

Patch	Reference OD	Measured OD	Deviation
1 (white)	0,10		
2	0,20		
3	0,31		
4	0,43		
5	0,56		
6	0,70		
7	0,87		
8	1,05		
9	1,27		
10	1,53		
11	1,86		
12 (black)	2,30		
BG	1,11		

\*Measured with X-Rite eXact

[https://www.image-engineering.de/content/products/charts/IE\\_reference\\_data\\_accuracy.pdf](https://www.image-engineering.de/content/products/charts/IE_reference_data_accuracy.pdf)

**Acceptance protocol**

SN:   
Contrast: 1000 :1  
Date:   
Operator:

**Optical density (OD) values gray patches\***

**Patch Reference OD Measured OD Deviation**

Patch	Reference OD	Measured OD	Deviation
1 (white)	0,11		
2	0,22		
3	0,34		
4	0,48		
5	0,63		
6	0,80		
7	0,99		
8	1,22		
9	1,49		
10	1,85		
11	2,33		
12 (black)	3,11		
BG	1,11		

\* Measured with X-Rite 361T and values adjusted to diffuse illumination

[https://www.image-engineering.de/content/products/charts/IE\\_reference\\_data\\_accuracy.pdf](https://www.image-engineering.de/content/products/charts/IE_reference_data_accuracy.pdf)

**Acceptance protocol**

SN: \_\_\_\_\_  
Contrast: 4000 :1  
Date: \_\_\_\_\_  
Operator: \_\_\_\_\_

**Optical density (OD) values gray patches\***

**Patch Reference OD Measured OD Deviation**

Patch	Reference OD	Measured OD	Deviation
1 (white)	0,11		
2	0,23		
3	0,36		
4	0,50		
5	0,66		
6	0,85		
7	1,07		
8	1,33		
9	1,65		
10	2,08		
11	2,74		
12 (black)	4,11		
BG	1,11		

\* Measured with X-Rite 361T and values adjusted to diffuse illumination

[https://www.image-engineering.de/content/products/charts/IE\\_reference\\_data\\_accuracy.pdf](https://www.image-engineering.de/content/products/charts/IE_reference_data_accuracy.pdf)